

**Konformitätserklärung
Declaration of Conformity**

Wir

We

**B. Braun Melsungen AG
Carl-Braun-Str. 1
34212 Melsungen
Deutschland/Germany
SRN DE-MF-000000201**erklären in eigener Verantwortung,
dass die Produkte**Contiplex® D,
Contiplex® D NRFit®**Atraumatische Punktionsnadel mit
Kunststoffkapillare zur Plexusanästhesie
Basis UDI-DI: 4039239000008532U
(Artikelnummern siehe Anlage I)Katheterset für kontinuierliche periphere
Nervenblockaden
Basis UDI-DI: 4039239000008522S
(Artikelnummern siehe Anlage I)mit den Anforderungen der Medizinprodukte
Verordnung (EU) 2017/745 übereinstimmen**Konformitätsbewertungsverfahren**
nach Anhang IX
der oben genannten Verordnung**Klassifizierung**
gemäß Anhang VIII der oben genannten
Verordnung
Klasse IIa**Benannte Stelle**
TÜV SÜD Product Service GmbH
Kennnummer 0123**Gültig bis**
gemäß gültigem EU Zertifikat
(Nr. G10 012974 0611)hereby declare in our own responsibility
that the products**Contiplex® D,
Contiplex® D NRFit®**Atraumatic puncture needle with plastic capillary
for plexus anesthesia
Basic UDI-DI: 4039239000008532U
(article numbers see attachment I)Continuous peripheral nerve block catheter set
Basic UDI-DI: 4039239000008522S
(Artikelnummern siehe Anlage I)are in conformity with the requirements of the
Medical Device Regulation (EU) 2017/745**Conformity Assessment Procedure**
according to annex IX
of the Regulation named above**Classification**
according to annex VIII of the Regulation named
above
Class IIa**Notified Body**
TÜV SÜD Product Service GmbH
Identification number 0123**Valid until**
according to our valid EU Certificate
(No. G10 012974 0611)

Anlage I / Attachment I**Basic UDI-DI 4039239000008532U**

Art.-Nr. / Art. No.	Produktname / Product name	Klasse / Class
4898205	Contiplex [®] D	Ila
4898211	Contiplex [®] D	Ila
4898235	Contiplex [®] D	Ila

Basic UDI-DI 4039239000008522S

Art.-Nr. / Art. No.	Produktname / Product name	Klasse / Class
4898323	Contiplex [®] D	Ila
4898323NR	Contiplex [®] D NRFit [®]	Ila
4898325	Contiplex [®] D	Ila
4898325NR	Contiplex [®] D NRFit [®]	Ila
4898305	Contiplex [®] D	Ila
4898305NR	Contiplex [®] D NRFit [®]	Ila
4894235NCN	Contiplex [®] D	Ila
4898335	Contiplex [®] D	Ila
4898335NR	Contiplex [®] D NRFit [®]	Ila
4898308	Contiplex [®] D	Ila
4895819NCN	Contiplex [®] D	Ila
4898311	Contiplex [®] D	Ila
4898311NR	Contiplex [®] D NRFit [®]	Ila
4894391NCN	Contiplex [®] D	Ila

Document amendment information

Version	Description of the changes
1.0	Initial Version under 2017/745 MDR

Title: Declaration of Conformity - 194-003-MDR - Contiplex D Initiator: Andreea ? Schade

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: Schade, Andreea (smerande)
Title: HC-RA Senior Manager Regulatory Affairs
Date: Wednesday, 22 May 2024, 08:59 W. Europe Daylight Time
Meaning: Document signed as Author

UserName: Seidel, Stefan (seidstde)
Title: Head of Regulatory Affairs CoE Infusion & Pain Therapy
Date: Wednesday, 22 May 2024, 09:42 W. Europe Daylight Time
Meaning: Approve Document

UserName: Brand, Thomas (brantode)
Title: HC-QM-DE08 Vice President QM for non-active Medical Devices
Date: Wednesday, 22 May 2024, 18:35 W. Europe Daylight Time
Meaning: Approve Document
